

## BIOEQUIVALENCE SUMMARY TABLES FOR METERED DOSE INHALER PRODUCTS

Please note that the tables listed in this document only include the bioequivalence summary tables related to the **in vitro** and **in vivo PD** studies recommended for metered dose inhaler products.

For the bioequivalence summary tables related to the **in vivo PK** BE tests, the applicant should refer to the Bioequivalence Summary Tables published on the Office of Generic Drugs website at <https://www.fda.gov/downloads/Drugs/DevelopmentApprovalProcess/HowDrugsareDevelopedandApproved/ApprovalApplications/AbbreviatedNewDrugApplicationANDAGenerics/UCM120957.pdf>

For the bioequivalence summary tables related to the **in vivo Clinical Endpoint** BE tests, the applicant should refer to the Bioequivalence Summary Tables published on the Office of Generic Drugs website at <https://www.fda.gov/downloads/Drugs/DevelopmentApprovalProcess/HowDrugsareDevelopedandApproved/ApprovalApplications/AbbreviatedNewDrugApplicationANDAGenerics/UCM400548.pdf>

**Table 1. Formulation Table**

INGREDIENTS	TEST		
	Amount per Actuation	Amount per mL	% (w/w)
<b>TOTALS</b>			
<b>NET FILL WEIGHT</b>			

**Table 2. Batch Information**

TEST								
Study Type	Lot No.	Potency***	Lot Size (# of Canisters)****		Manufacture Date for Test Expiration Date for Reference	API Lot(s)	Critical Excipient (e.g. surfactant, co-solvent, etc) Lot (s)	Container Closure System (e.g. Valve, Actuator, Canister) lot(s)
			Theoretical	Actual				
Bioequivalence study (PK study) *								
Bioequivalence study (PD study) *								
<i>In-Vitro</i> equivalence studies **								
REFERENCE								
Bioequivalence study (PK study) *								
Bioequivalence study (PD study) *								
<i>In-Vitro</i> equivalence studies **								

\* If recommended

\*\* Include lot numbers from each *in vitro* test

\*\*\* Data obtained from Certificate of Analysis

\*\*\*\* The size of exhibited batches should be at least one-third of the to-be-marketed production batch size

**Table 3. Device Comparability**

	<b>TEST</b>	<b>REFERENCE</b>
<b>Canister</b>		
<b>Canister Supplier</b>		
<b>Material</b>		
<b>Canister Volume</b>		
<b>Valve</b>		
<b>Valve Supplier</b>		
<b>Metering Volume</b>		
<b>Gasket and Seat Elastomers (material)</b>		
<b>Metering Chamber and Body (material)</b>		
<b>Core and Core Extension/Base</b>		
<b>Actuator</b>		
<b>Actuator Supplier</b>		
<b>Actuator Orifice Diameter (µm)</b>		
<b>Material</b>		
<b>Protection Cap Description</b>		
<b>Dose Counter/Indicator</b>	<input type="checkbox"/> Yes <input type="checkbox"/> No	<input type="checkbox"/> Yes <input type="checkbox"/> No
<b>Number of Doses</b>		
<b>Cleaning instructions (similar cleaning instruction and frequency? *)</b>		

\*With alternate device design, the applicant should provide justification and evidence to support that there will be no confusion with respect to cleaning.

**Table 4. Actuation Methods**

<p><b>Which tests (if any) used MANUAL actuation?</b></p>				
<p><b>If some tests used manual actuation(s), describe methods used to avoid Test to RLD bias in dose release.</b></p>				
<p><b>Which tests (if any) used AUTOMATED actuation?</b></p>				
<p><b>What were the parameters of automated actuation? (units)*</b></p>			<p>Test</p>	<p>RLD</p>
	<p><b>Force Driven System [e.g., MDI AS, Hand Actuation Monitor (HAM), etc.]</b></p>	<p><b>Force (kg or N)</b></p>		
		<p><b>Force Rise Time (msec)</b></p>		
		<p><b>Force Fall time (msec)</b></p>		
		<p><b>Hold Time (msec)</b></p>		
		<p><b>Agitation Shaking (msec)</b></p>		
	<p><b>Velocity Driven Actuator [e.g., Vereo Actuator SFMDx, SPRAYTEC with SPRAYER module, etc.]</b></p>	<p><b>Velocity (mm/s)</b></p>		
		<p><b>Acceleration (mm/s<sup>2</sup>)</b></p>		
		<p><b>Initial Hold Time (msec)</b></p>		
		<p><b>Hold Time (msec)</b></p>		
		<p><b>Final Delay (msec)</b></p>		
		<p><b>Pre-Stage Position</b></p>		
<p><b>Are the actuation parameters the same for the test and reference products? If No, please comment</b></p>	<p><input type="checkbox"/> Yes <input type="checkbox"/> No</p>			

\*Parameters may vary depending on the instrument

**The Table 5 Series is for Single Actuation Content through Container Life Test**

**Table 5. 1. Study Information**

<b>Study No.</b>	
<b>Study Site Name and address</b>	
<b>Principal Investigator</b>	
<b>Study Dates</b>	
<b>SOP No.</b>	
<b>SOP Effective Date</b>	
<b>SOP Title</b>	
<b>Test Method Description</b>	
<b>Testing Equipment Used</b> (e.g., name, model, etc.)	
<b>Operating Conditions for Testing Equipment Used</b> (e.g., temperature, humidity, etc.)	
<b>Analytical Method Description</b>	
<b>Analytical Equipment Used</b> (e.g., name, model, etc.)	

**Table 5. 2. Analytical Method Validation for HPLC**

<b>Information Requested</b>	
<b>Analytical method validation report location</b>	Provide the volume(s) and page(s)
<b>Analyte</b>	Provide the name(s) of the analyte(s)
<b>Internal Standard (IS)</b>	Only if applicable
<b>Method description</b>	Brief descriptions of extraction method; analytical
<b>Selectivity or Specificity</b>	Brief comments
<b>Limit of quantitation</b>	LOQ, unit
<b>Detection Limit</b>	LOD, unit
<b>Linearity Range (ng, mcg/mL)</b>	Range, unit
<b>Linearity (R<sup>2</sup>) (e.g., 0.99)</b>	
<b>Accuracy (% recovery)</b>	Avg.: HQC: MQC: LQC:
<b>Precision – Repeatability (CV%)</b>	QC
<b>Precision -- Intermediate Precision</b>	By Date:  By Analyst:
<b>Bench-top stability (hrs (CV%)) (working std solution) (e.g. 2 days @ room temperature)</b>	
<b>Stock solution stability (days (CV %))</b>	Only if applicable
<b>Robustness</b>	Brief comments

**Calibration of Manual and/or Automated MDI Actuation (For Single Actuation Content)**

**Table 5.3. Precision and Ruggedness**

	Precision	Ruggedness	
<b>Content assay (µg) (Mean and CV%)</b>		Day 1*:	Day 2*:
		Analyst 1:	Analyst 2:
		Unit 1**:	Unit 2**
<b>Shot weight (mg) (Mean and CV%)</b>		Day 1*:	Day 2*:
		Analyst 1:	Analyst 2:
		Unit 1**:	Unit 2**:
<b>% Difference in Content assay means</b>		Between Day 1 and 2:	
		Between Analyst 1 and 2:	
		Between Unit 1 and 2:	
<b>Content assay (% CV)</b>		Inter day:	
		Inter analyst	
<b>% Difference in shot weight means</b>		Inter unit	
		Between Day 1 and 2:	
		Between Analyst 1 and 2:	
<b>Shot weight (% CV)</b>		Between Unit 1 and 2:	
		Inter day:	
		Inter analyst:	
<b>Acceptance criteria defined by SOP</b>		Inter unit:	
		<i>Example</i>	
		Precision: Intermediate Precision by Date: Intermediate Precision by Analyst: Intermediate Precision by Unit: % Difference Day-to-Day: % Difference Analyst-to-Analyst: % Difference Unit-to-Unit:	
<b>Reference Product lot numbers, expiration dates</b>			
<b>Number of units</b>			
<b>Number of sprays/unit</b>			
<b>Automated or manual actuation used</b>	Automated / Manual		

\* Ruggedness by day: By same analyst

\*\* Ruggedness by units: If more than 1 unit used in the validation

**Table 5. 4. Results Summary – Single Actuation Content**

SINGLE ACTUATION CONTENT THROUGH CONTAINER LIFE												
	Spray #	Mean				Variability (%CV)					Mean Ratio (T/R)	
		Drug Mass (mg)		% label claim		Within Lot (n=10)			Between Lot (n=3)	Total (n=30)	Arith (n=30)	Geo (n=30)
		Arith	Geo	Arith	Geo	Lot 1	Lot 2	Lot 3				
BEG	Test											
	Ref											
MID	Test											
	Ref											
END	Test											
	Ref											

**Table 5.4.1. Summary of Population Bioequivalence Results**

Variable	Mean (log Scale)		Mean Difference (log Scale)	Standard Deviation		Sigma T /Sigma R Ratio
	Test	Reference		Sigma T	Sigma R	
Scaled		Linearized Point Estimate		95% Upper Confidence Bound		Pass or Fail PBE
Reference-scaled						
Constant-scaled						

The Single Actuation Content comparison of the T and R products is based on the population bioequivalence (PBE). Refer to draft budesonide inhalation suspension BE guidance for additional information regarding PBE analysis procedures.

(<http://www.fda.gov/downloads/Drugs/GuidanceComplianceRegulatoryInformation/Guidances/UCM319977.pdf>).



**The Table 6 Series is for Priming & Re-priming Test**

**Table 6.1. Study Information**

<b>Study No.</b>	
<b>Study Site Name and Address</b>	
<b>Principal Investigator</b>	
<b>Study dates</b>	
<b>SOP No.</b>	
<b>SOP Effective Date</b>	
<b>SOP Title</b>	
<b>Test Method Description</b>	
<b>Testing Equipment Used</b> (e.g., name, model, etc)	
<b>Operating Conditions for Testing Equipment Used</b> (e.g., temperature, humidity, etc..)	
<b>Analytical Method Description</b>	
<b>Analytical Equipment Used</b> (e.g., name, model, etc)	

**Note:** The repriming test should be performed following storage for the specified period of non-use after initial use and/or other conditions (e.g., dropping), if the reference product labeling provides such repriming information.

**Table 6. 2. Analytical Method Validation for HPLC**

To be completed only if different from Table 5.2

<b>Information Requested</b>	
<b>Analytical method validation report location</b>	Provide the volume(s) and page(s)
<b>Analyte</b>	Provide the name(s) of the analyte(s)
<b>Internal Standard (IS)</b>	Only if applicable
<b>Method description</b>	Brief descriptions of extraction method; analytical
<b>Selectivity or Specificity</b>	Brief comments
<b>Limit of quantitation</b>	LOQ, unit
<b>Detection Limit</b>	LOD, unit
<b>Linearity Range (ng, mcg/mL)</b>	Range, unit
<b>Linearity (R<sup>2</sup>) (e.g., 0.99)</b>	
<b>Accuracy (% recovery at the high and low concentrations)</b>	Avg.: HQC: MQC: LQC:
<b>Precision – Repeatability (CV%)</b>	QC
<b>Precision -- Intermediate Precision</b>	By Date:  By Analyst:
<b>Bench-top stability (hrs (CV%)) (working std solution) (e.g. 2 days @ room temperature)</b>	
<b>Stock solution stability (days (CV %))</b>	Only if applicable
<b>Robustness</b>	Brief comments

**Table 6.3. Precision and Ruggedness**

To be completed only if different from Table 5.3

	Precision	Ruggedness	
Content assay (µg) (Mean and CV%)		Day 1*:	Day 2*:
		Analyst 1:	Analyst 2:
		Unit 1**:	Unit 2**
Shot weight (mg) (Mean and CV%)		Day 1*:	Day 2*:
		Analyst 1:	Analyst 2:
		Unit 1**:	Unit 2**:
%Difference in content assay means		Between Day 1 and 2:	Inter day %CV
		Between Analyst 1 and 2:	Inter analyst %CV
		Between Unit 1 and 2:	Inter unit %CV
%Difference in shot weight means		Between Day 1 and 2:	Inter day %CV
		Between Analyst 1 and 2:	Inter analyst %CV
		Between Unit 1 and 2:	Inter unit %CV
Acceptance criteria defined by SOP	<i>Example</i> Precision: Intermediate Precision by Date: Intermediate Precision by Analyst: Intermediate Precision by Unit: % Difference Day-to-Day: % Difference Analyst-to-Analyst: % Difference Unit-to-Unit:		
RLD lot numbers, expiration dates			
Number of units			
Number of sprays/unit			
Automated or manual actuation used	Automated / Manual		

\* Ruggedness by day: By same analyst

\*\* Ruggedness by units: If more than 1 unit used in the validation



**Table 6. 4. 1. Summary of Population Bioequivalence Results**

Variable	Mean (log Scale)		Mean Difference (log Scale)	Standard Deviation		Sigma T /Sigma R Ratio
	Test	Reference		Sigma T	Sigma R	
<b>Priming</b>						
Scaled		Linearized Point Estimate	95% Upper Confidence Bound	Pass or Fail PBE		
Reference-scaled						
Constant-scaled						

Variable	Mean (log Scale)		Mean Difference (log Scale)	Standard Deviation		Sigma T /Sigma R Ratio
	Test	Reference		Sigma T	Sigma R	
<b>Repriming</b>						
Scaled		Linearized Point Estimate	95% Upper Confidence Bound	Pass or Fail PBE		
Reference-scaled						
Constant-scaled						

**The Table 7 Series is for Aerodynamic Particle Size Distribution (APSD) by Cascade Impaction**

**Table 7. 1. Study Information**

<b>Study No.</b>	
<b>Study Site Name and address</b>	
<b>Principal Investigator</b>	
<b>Study dates</b>	
<b>SOP No.</b>	
<b>SOP Effective Date</b>	
<b>SOP Title</b>	
<b>Testing Method Description</b> [Eg. Test batches, B and E Lifestages, Number of canisters/batch, CI set up, flow rate determination, plate/cup coating, priming regimen, actuation method, filter, extraction diluent]	
<b>Testing Equipment Used</b> [e.g., name, model, etc, equipment includes but not limited to USP Apparatus (ACI or NGI), Flow Controller, Flow meter, Pump]	
<b>Operating Conditions for Testing Equipment Used</b> (e.g., temperature, humidity, etc..)	
<b>Analytical Method Description</b>	
<b>Analytical Equipment Used</b> (e.g., name, model, etc)	

## Validation Summary Tables for Aerodynamic Particle Size Distribution (APSD) by Cascade Impaction

**Table 7.2. Analytical Method Validation for HPLC**

<b>Information Requested</b>	
<b>Analytical method validation report location</b>	Provide the volume(s) and page(s)
<b>Analyte</b>	Provide the name(s) of the analyte(s)
<b>Internal Standard (IS) (If applicable)</b>	
<b>Method description</b>	Brief descriptions of extraction method; analytical
<b>Selectivity or Specificity</b>	
<b>Limit of quantitation (unit)</b>	LOQ, unit
<b>Detection Limit (unit)</b>	LOD, unit
<b>Linearity Range (ng, mcg/mL)</b>	Range, unit
<b>Linearity (R<sup>2</sup>) (e.g., 0.99)</b>	
<b>Accuracy (% recovery)</b>	Avg.: HQC: MQC: LQC:
<b>Precision – Repeatability, (%CV)</b>	QC
<b>Intermediate Precision</b>	<u>By Date:</u>  <u>By Analyst:</u>
<b>Bench-top stability (hrs (CV%)) (working std solution) (e.g. 2 days @ room temperature)</b>	
<b>Stock solution stability (days (CV %)) (If applicable)</b>	
<b>Robustness</b>	

**Table 7.3 Method Validation for Cascade Impaction**

	Precision (n = #)	Robustness (By Analyst/Day)	
Impactor-Sized Mass, (µg, mg/actuation) (mean and CV%)		Analyst 1	Analyst 2
		Day 1	Day 2
Fine Particle Mass (µg, mg/actuation) (mean and CV%)		Analyst 1	Analyst 2
		Day 1	Day 2
Mass Median Aerodynamic Diameter (µm) (mean and %CV)		Analyst 1	Analyst 2
		Day 1	Day 2
Geometric Standard Deviation (mean and %CV, if applicable)		Analyst 1	Analyst 2
		Day 1	Day 2
Delivered Dose (µg, mg) (mean and CV%)		Analyst 1:	Analyst 2:
		Day 1	Day 2
Mass Balance (mg) (mean and CV%)		Analyst 1:	Analyst 2:
		Day 1	Day 2
% Difference in Impactor-Sized Mass		Between Analyst 1 and 2:	
		Between Day 1 and 2:	
Impactor-Sized mass (% CV)		Inter analyst:	
		Inter day:	
% Difference in Fine Particle Mass		Between Analyst 1 and 2:	
		Between Day 1 and 2:	
Fine Particle Mass (% CV)		Inter analyst:	
		Inter day:	
% Difference in Delivered Dose		Between Analyst 1 and 2:	
		Between Day 1 and 2:	
Delivered Dose (% CV)		Inter analyst:	
		Inter day:	
%Difference in Mass Balance		Between Analyst 1 and 2:	
		Between Day 1 and 2:	
Mass Balance (% CV)		Inter analyst:	
		Inter day:	
Acceptance criteria defined by SOP	Example Precision: Intermediate Precision by Analyst: Intermediate Precision by Day: % Difference Analyst-to-Analyst: % Difference Day-to-Day:		
Reference Product lot numbers			
Number of units			



<b>Number of actuation/unit</b>	
<b>Automated or manual actuation used</b>	Automated / Manual

**Table 7.4 Results Summary – Aerodynamic Particle Size Distribution by Cascade Impaction**

<b>Aerodynamic Particle Size Distribution</b>										
		<b>Mean Drug Deposition (µg)</b>		<b>Variability (%CV)</b>					<b>Mean Ratio (T/R)</b>	
				Within Lot (n=20)			Between Lot (n=3)	Total (n=60)	Arith (n=60)	Geo (n=60)
		Arith	Geo	Lot 1	Lot 2	Lot 3				
<b>Delivered Dose (µg)</b>	<b>Test</b>									
	<b>Ref</b>									
<b>Fine Particle Mass (µg)</b>	<b>Test</b>									
	<b>Ref</b>									
<b>Impactor-Sized Mass (µg)</b>	<b>Test</b>									
	<b>Ref</b>									

Test Lot #1 – xxxxx, Lot #2 – xxxxx, Lot #3 – xxxxx; Reference Lot #1 – xxxxx, Lot #2 – xxxxx, Lot #3 – xxxxx

<b>MASS BALANCE* (% of label claim)</b>		
		<b>Arithmetic Mean and Range (Min – Max) (n=30)</b>
<b>Mass Balance (%)</b>	<b>Test</b>	
	<b>Ref</b>	

\* Determined from mouthpiece adapter, the induction port, each stage of the cascade impactor (CI) and the filter, and any other accessories.

**Table 7.5 Summary of Population Bioequivalence Results**

Variable	Mean (log Scale)		Mean Difference (log Scale)	Standard Deviation		Sigma T/Sigma R Ratio
	Test	Reference		Sigma T	Sigma R	
ISM						
Scaled	Linearized Point Estimate		95% Upper Confidence Bound		Pass or Fail PBE	
Reference-scaled						
Constant-scaled						

**The Table 8 Series is for Spray Pattern Test**

**Table 8.1 Study Information**

<b>Study No.</b>	
<b>Study Site Name and Address</b>	
<b>Principal Investigator</b>	
<b>Study dates</b>	
<b>SOP No.</b>	
<b>SOP Effective Date</b>	
<b>SOP Title</b>	
<b>Testing Method Description</b>	
<b>Testing Equipment Used</b> (e.g., name, model, etc)	
<b>Image Analysis Apparatus Used</b> (i.e., automated = Laser Imaging; or manual = TLC)	
<b>Operating Conditions for Testing Equipment Used</b> (e.g., temperature, humidity, etc..)	

## Validation Summary Table for Spray Pattern

**Table 8.2 Precision and Ruggedness**

	Distance (e.g., 3 cm and 6 cm)	Precision	Ruggedness	
Area <sup>1</sup> (mean and CV%)	Distance 1		Day 1*:	Day 2*:
			Analyst 1:	Analyst 2:
	Distance 2		Day 1*:	Day 2*:
			Analyst 1:	Analyst 2:
Ovality Ratio (mean and CV%)	Distance 1		Day 1*:	Day 2*:
			Analyst 1:	Analyst 2:
	Distance 2		Day 1*:	Day 2*:
			Analyst 1:	Analyst 2:
Difference in Area <sup>1</sup> (%)	Distance 1		Between Day 1 and 2:	
			Between Analyst 1 and 2:	
	Distance 2		Between Day 1 and 2:	
			Between Analyst 1 and 2:	
Area (% CV)	Distance 1		Inter day:	
			Inter analyst:	
	Distance 2		Inter day:	
			Inter analyst:	
Difference in Ovality Ratio (%)	Distance 1		Between Day 1 and 2:	
			Between Analyst 1 and 2:	
	Distance 2		Between Day 1 and 2:	
			Between Analyst 1 and 2:	
Ovality Ratio (% CV)	Distance 1		Inter day:	
			Inter analyst:	
	Distance 2		Inter day:	
			Inter analyst:	
<b>Acceptance criteria defined by SOP</b>		<i>Example</i> Precision: Intermediate Precision by Date: Intermediate Precision by Analyst: % Difference Day-to-Day: % Difference Analyst-to-Analyst:		
<b>Reference Product lot numbers</b>				
<b>Number of units</b>				
<b>Number of sprays/unit</b>				
<b>Automated or manual actuation used</b>				

\* Ruggedness by day: By same analyst

1. This parameter varies with the type of spray pattern analysis. If it is an automated analysis, e.g., Laser imaging, "area" should be used. If it is a manual analysis, e.g., TLC, "Dmax" should be used.



**Table 8.3.1. Summary of Population Bioequivalence Results**

Variable	Mean (log Scale)		Mean Difference (log Scale)	Standard Deviation		Sigma T /Sigma R Ratio
	Test	Reference		Sigma T	Sigma R	
Area at X cm						
Scaled		Linearized Point Estimate	95% Upper Confidence Bound	Pass or Fail PBE		
Reference-scaled						
Constant-scaled						

Variable	Mean (log Scale)		Mean Difference (log Scale)	Standard Deviation		Sigma T /Sigma R Ratio
	Test	Reference		Sigma T	Sigma R	
Area at Y cm						
Scaled		Linearized Point Estimate	95% Upper Confidence Bound	Pass or Fail PBE		
Reference-scaled						
Constant-scaled						

Variable	Mean (log Scale)		Mean Difference (log Scale)	Standard Deviation		Sigma T /Sigma R Ratio
	Test	Reference		Sigma T	Sigma R	
Ovality Ratio at X cm						
Scaled		Linearized Point Estimate	95% Upper Confidence Bound	Pass or Fail PBE		
Reference-scaled						
Constant-scaled						

Variable	Mean (log Scale)		Mean Difference (log Scale)	Standard Deviation		Sigma T /Sigma R Ratio
	Test	Reference		Sigma T	Sigma R	
Ovality Ratio at Y cm						
Scaled		Linearized Point Estimate	95% Upper Confidence Bound	Pass or Fail PBE		
Reference-scaled						
Constant-scaled						

**The Table 9 Series is for Plume Geometry Test**

**Table 9.1. Study Information**

<b>Study No.</b>	
<b>Study Site Name and Address</b>	
<b>Principal Investigator</b>	
<b>Study dates</b>	
<b>SOP No.</b>	
<b>SOP Effective Date</b>	
<b>SOP Title</b>	
<b>Testing Method Description</b> (e.g., Actuation distance; criteria for defining the plume angle and width, etc.)	
<b>Criteria for defining plume angle and width borders</b>	
<b>Testing Equipment Used</b> (e.g., name, model, etc)	
<b>Image Analysis Apparatus Used</b>	
<b>Operating Conditions for Testing Equipment Used</b> (e.g., temperature, humidity, etc..)	

The applicant needs to submit representative photographs (manual) or digital images (automated) and spray intensity (actuation) profiles as supportive data.

## Validation Summary Table for Plume Geometry

**Table 9.2 Precision and Ruggedness**

	Precision	Ruggedness	
Plume Width (mean and CV%)		Day 1*:	Day 2*:
		Analyst 1:	Analyst 2:
Plume Angle (mean and CV%)		Day 1*:	Day 2*:
		Analyst 1:	Analyst 2:
Difference in Plume Width (%)		Between Day 1 and 2:	
		Between Analyst 1 and 2:	
Plume Width (% CV)		Inter day:	
		Inter analyst:	
Difference in Plume Angle (%)		Between Day 1 and 2:	
		Between Analyst 1 and 2:	
Plume Angle (% CV)		Inter day:	
		Inter analyst:	
Acceptance criteria defined by SOP	<i>Example</i> Precision: Intermediate Precision by Date: Intermediate Precision by Analyst: % Difference Day-to-Day: % Difference Analyst-to-Analyst:		
Reference Product lot numbers			
Number of units			
Number of sprays/unit			
Automated or manual actuation used			

\*Ruggedness by day: By same analyst





For SAS Data Tables for MDI product In Vitro Bioequivalence Study Data Submission, Please Refer to the related section in “Bioequivalence Summary Tables for Aqueous Nasal Spray Products” published on the Office of Generic Drugs at <https://www.fda.gov/downloads/Drugs/DevelopmentApprovalProcess/HowDrugsareDevelopedandApproved/ApprovalApplications/AbbreviatedNewDrugApplicationANDAGenerics/UCM209446.pdf>

The Table 10 Series is for the Pharmacodynamic (PD) Bioequivalence (BE) Bronchoprovocation Study

**Table 10.1. Study Information**

<b>Study Number</b>	
<b>Study Title</b>	
<b>Clinical Site(s) (Name &amp; Address)</b>	
<b>Principal Clinical Investigator(s)</b>	
<b>Clinical Study Date Range</b>	

**Table 10.2. Product Information**

<b>Product</b>	<b>Test</b>	<b>Reference</b>	<b>Placebo</b>
<b>Product Name</b>			
<b>Manufacturer</b>			
<b>Batch/Lot No.</b>			
<b>Manufacture Date</b>			
<b>Expiration Date</b>			
<b>Strength</b>			
<b>Bio-batch Size</b>			
<b>Production Batch Size</b>			
<b>Dosage Form</b>			
<b>Potency, %</b>			
<b>Content Uniformity (Mean, %CV)</b>			
<b>Dose Administered</b>			
<b>Route of Administration</b>			

**Table 10.3. Demographic Profile of Subjects Completing the BE Study**

		PD Study No.		
		Treatment Groups		
		Test Product N =	Reference Product N =	Placebo N =
<b>Age (years)</b>	<b>Mean ± SD</b> <b>Range</b>			
<b>Age Groups</b>	<b>&lt; 18</b>	N (%)	N (%)	N (%)
	<b>18 – 40</b>	N (%)	N (%)	N (%)
	<b>41 – 64</b>	N (%)	N (%)	N (%)
	<b>65 – 75</b>	N (%)	N (%)	N (%)
	<b>&gt; 75</b>	N (%)	N (%)	N (%)
<b>Sex</b>	<b>Male</b>	N (%)	N (%)	N (%)
	<b>Female</b>	N (%)	N (%)	N (%)
<b>Race</b>	<b>Asian</b>	N (%)	N (%)	N (%)
	<b>Black</b>	N (%)	N (%)	N (%)
	<b>Caucasian</b>	N (%)	N (%)	N (%)
	<b>Hispanic</b>	N (%)	N (%)	N (%)
	<b>Other</b>	N (%)	N (%)	N (%)
<b>BMI</b>	<b>Mean ± SD</b> <b>Range</b>			
<b>Other Factors</b>				

**Table 10.4. Dropout Information**

Subject No.	Reason for dropout/replacement	Period	Replaced?	Replaced With

**Table 10.5. Incidence of Adverse Events in the PD BE Study**

Body System / Adverse Event	Reported Incidence by Treatment Groups		
	PD Study No.		
	Test	Reference*	Placebo
Body as a whole			
Dizziness	N (%)	N (%)	N (%)
Etc.	N (%)	N (%)	N (%)
Cardiovascular			
Hypotension	N (%)	N (%)	N (%)
Etc.	N (%)	N (%)	
Gastrointestinal			
Emesis*	N (%)	N (%)	N (%)
Constipation	N (%)	N (%)	N (%)
Etc.	N (%)	N (%)	N (%)
Other organ sys.			
	N (%)	N (%)	N (%)
	N (%)	N (%)	N (%)
Total	N (%)	N (%)	N (%)

\*Please separate the R treatments by dose

**Table 10.6. Protocol Deviations**

Type	Subjects with deviation				
	Test	Reference Dose 1	Reference Dose 2	Placebo	Total

**Table 10.7. Statistical Summary for the PD BE Study**

**Table 10.7a. Point Estimates and 90% Confidence Intervals, Raw Data**

Drug name Dose Pharmacodynamic Study No. (study number), N=N1 Point Estimates and 90% Confidence Intervals		
Parameter	Point Estimate	90% C.I.
F		

**Note:** Please submit the estimated value for E0, ED50R, and EmaxR. ED50R and EmaxR refer to the modeled ED50 and Emax for the reference product only.

**Table 10.7b. Point Estimates and 90% Confidence Intervals, Bootstrapping Procedure**

Drug name Dose Pharmacodynamic Study No. (study number), N=N1 Point Estimates and 90% Confidence Intervals		
Parameter	Point Estimate	90% C.I.
F		

**Note:** Please submit the estimated value for E0, ED50R, and EmaxR. ED50R and EmaxR refer to the modeled ED50 and Emax for the reference product only.

**Table 10.8. PD BE Study, Additional Information**

Subjects excluded in statistical analysis for each period and reason for exclusion [include Subject #, Product (T, R, or placebo), and Dose]	
Subjects that failed to reduce the FEV1 by 20% following the highest dose of methacholine [include Subject #, Product (T, R, or placebo), and Dose]	
Stepwise method of PC20 estimation	

**Table 10.9. SAS Data Table for MDI product In Vivo PD BE Study Data Submission**

Data in this table should be arranged in columns as shown in examples. Data sets should be submitted as SAS Transport files.

<b>Variable Name</b>	<b>Variable Type</b>	<b>Content</b>	<b>Notes</b>
Subject ID	Numeric	Numeric values	Identifier for subject
Treatment	Numeric	Numeric values	Identifier for treatment (product and dose)
Period	Numeric	Numeric values	Identifier for period
Dose	Numeric	Numeric values	Identifier for dose
PC20	Numeric	Numeric values	The provocative concentration or dose, respectively, of the methacholine challenge agent required to reduce the forced expiratory volume in one second (FEV1) by 20% following administration of differing doses of study drug (or placebo) by inhalation.

<b>Subject ID</b>	<b>Treatment</b>	<b>Period</b>	<b>Dose</b>	<b>PC20</b>	<b>lnPC20</b>