

Welcome to today's FDA/CDRH Webinar

*Thank you for your patience while additional
time is provided for participants to join the call.*

**Please connect to the audio portion of the
webinar now:**

U.S. Callers Dial: 888-455-1392

International Callers Dial: 1-773-799-3847

Passcode: 3598011

Transitioning from the Use of Decontaminated Disposable Respirators

Office of Strategic Partnerships and Technology Innovation (OST)
Center for Devices and Radiological Health (CDRH)
U.S. Food and Drug Administration (FDA)

April 27, 2021

Agenda

- Overview of FDA Letter to Health Care Personnel & Facilities
 - Background
 - NIOSH-Approved Respirators Authorized under Emergency Use Authorization (EUA)
 - Recommendations
 - FDA Actions
- Resources
- Questions

Background

- On April 9, 2021, the FDA issued a letter to health care personnel and facilities recommending health care personnel and facilities **transition away from crisis capacity conservation strategies**, such as decontaminating or bioburden reducing disposable respirators for reuse.
- Based on the **increased domestic supply of new respirators** approved by the Centers for Disease Control and Prevention's (CDC) National Institute for Occupational Safety and Health (NIOSH) currently available to facilitate this transition, the FDA and CDC believe there is adequate supply of respirators to transition away from use of decontamination and bioburden reduction systems.

NIOSH-Approved Respirators Authorized under EUA

- During the COVID-19 public health emergency, **NIOSH-approved respirators, including N95 respirators, are authorized*** on a continual basis under the FDA emergency use authorization (EUA) for NIOSH-Approved air purifying respirators (includes single-use respirators and those designed to be reusable)
- There are **over 6,400 total respirator models or configurations on the NIOSH certified equipment list** which meet the NIOSH-Approved EUA criteria and thus are FDA-authorized, including:
 - **Over 600 filtering facepiece respirators (FFR) models** (of which there are over 530 N95 FFR models)
 - **Over 5,500 elastomeric respirator configurations**, including new elastomeric respirators without an exhalation valve
 - **Over 360 powered air-purifying respirators (PAPR) configurations**

* until the U.S. Department of Health and Human Services (HHS) Secretary's declaration that circumstances exist justifying authorization is terminated or the EUA is revoked

Recommendations

- **The FDA recommends** that health care personnel and facilities:
 - **Limit decontamination of disposable respirators.** Decontaminated respirators and respirators that have undergone bioburden reduction should be used only when there are insufficient supplies of new FFRs or if you are unable to obtain any new respirators.
 - **Transition away from a crisis capacity strategy for respirators**, such as decontamination of N95 and other FFRs.
 - **Increase inventory of available NIOSH-approved respirators**—including **N95s and other FFRs, elastomeric respirators**, including new elastomeric respirators without an exhalation valve that can be used in the operating room, and **powered air-purifying respirators (PAPRs)**. Even if you are unable to obtain the respirator model that you would prefer, the FDA recommends that you obtain and use a new respirator before decontaminating or bioburden reducing a preferred disposable respirator.

FDA Actions

- The FDA **will continue to monitor supply and demand** to assess respirator availability as facilities systematically transition away from the most extreme measures of respirator conservation (that is, crisis capacity strategies) to contingency and eventually conventional use.
- Respirators, specifically surgical respirators, presently **remain on the FDA's device shortage list**.
- The FDA **will continue to keep health care personnel and the public informed** if new or additional information becomes available.

Resources

- FDA Recommends Transition from Use of Decontaminated Disposable Respirators - Letter to Health Care Personnel and Facilities:
<https://www.fda.gov/medical-devices/letters-health-care-providers/fda-recommends-transition-use-decontaminated-disposable-respirators-letter-health-care-personnel-and>
- FDA Considerations for Selecting Respirators for Your Health Care Facility
<https://www.fda.gov/medical-devices/coronavirus-covid-19-and-medical-devices/considerations-selecting-respirators-your-health-care-facility>
- CDC Strategies for Optimizing the Supply of N95 Respirators:
<https://www.cdc.gov/coronavirus/2019-ncov/hcp/respirators-strategy/index.html>
- CDC Personal Protective Equipment (PPE) Burn Rate Calculator:
<https://www.cdc.gov/coronavirus/2019-ncov/hcp/ppe-strategy/burn-calculator.html>
- OSHA Protecting Workers: Guidance on Mitigating and Preventing the Spread of COVID-19 in the Workplace:
<https://www.osha.gov/coronavirus/safework>
- OSHA enforcement memoranda that include time-limited discretions:
https://www.osha.gov/coronavirus/standards#temp_enforcement_guidance

Questions?

Email: CDRH-COVID19-SurgicalMasks@fda.hhs.gov

Slide Presentation, Transcript, and Webinar Recording
will be available at:

www.fda.gov/training/cdrhlearn

Under Heading: Specialty Technical Topics and
Sub-heading Personal Protective Equipment (PPE)

Please complete a short survey about your FDA CDRH webinar experience.

The survey can be found at www.fda.gov/CDRHWebinar

immediately following the conclusion of the live webinar.