



March 25, 2022

Cameron Ball, PhD  
Chief Executive Officer  
Uh-Oh Labs Inc.  
3485 Victor St  
Santa Clara, CA 95054

Re: EUA210666/S001  
Trade/Device Name: UOL COVID-19 Test  
Dated: February 24, 2022  
Received: February 24, 2022

Dear Dr. Ball:

This is to notify you that your request to update the shelf-life of the UOL COVID-19 Test to 8 months when stored at 2–30°C in accordance with the storage conditions on the different components, is granted. Upon review, we concur that the data and information submitted in EUA210666/S001 supports the requested updates for use with the UOL COVID-19 Test. We also concur with the minor updates to the UOL COVID-19 Test Instructions for Use to correct grammatical errors and changes for accuracy and ease of reading. By submitting this EUA revision for review by the Food and Drug Administration (FDA), you have complied with the Conditions of Authorization stated in the letter authorizing the emergency use of the UOL COVID-19 Test issued on February 8, 2022.

Sincerely yours,

---

Uwe Scherf, M.Sc., Ph.D.  
Director, Division of Microbiology Devices  
OHT7: Office of In Vitro Diagnostics and Radiological Health  
Office of Product Evaluation and Quality  
Center for Devices and Radiological Health