



October 26, 2022

Khushvanreep Singh M.S.  
Regulatory Affairs Specialist  
Roche Molecular Systems, Inc.  
4300 Hacienda Drive  
Pleasanton, CA 94588

Re: EUA210388/S002/A001  
Trade/Device Name: cobas SARS-CoV-2 Nucleic acid test for use on the cobas Liat System  
Dated: July 12, 2022 and August 4, 2022  
Received: July 13, 2022 and August 4, 2022

Dear Khushvanreep Singh:

This is to notify you that your request to extend the shelf life for the cobas SARS-CoV-2 Quality Control Kit reagent to 18 months at 2-8°C, based on the results of a real-time stability study, is granted. Upon review, we concur that the data and information submitted in EUA210388/S002/A001 support the requested updates for the cobas SARS-CoV-2 Nucleic acid test for use on the cobas Liat System. By submitting this EUA revision for review by the Food and Drug Administration (FDA), you have complied with the Conditions of Authorization stated in the letter authorizing the emergency use of the cobas SARS-CoV-2 Nucleic acid test for use on the cobas Liat System issued on June 17, 2021.

Sincerely yours,

---

Uwe Scherf, M.Sc., Ph.D.  
Director, Division of Microbiology Devices  
OHT7: Office of In Vitro Diagnostics  
Office of Product Evaluation and Quality  
Center for Devices and Radiological Health